


長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-310; Serial No.: 1401 2233; Channel No.: 1

FREEZER TEMPERATURE RANGE : -15~-25°C

DATE: 29, Dec 2021 ~ 25, Jan 2022



統計表

| #1 | |
|--------|---------|
| 最小 T | -23.45 |
| 最小 t | -23.45 |
| 最大值 T | -18.72 |
| 最大值 t | -18.72 |
| 平均值 T | -20.927 |
| 平均值 t | -20.927 |
| 標準偏差 T | 0.821 |
| 標準偏差 t | 0.821 |
| MKT T | -20.929 |
| MKT t | -20.929 |

T: 整個週期
t: 顯示週期

Comment:


長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-310; Serial No.: 1401 2233; Channel No.: 1

FREEZER TEMPERATURE RANGE : -15~-25°C

DATE: 25, Jan 2022 ~ 24, Feb 2022



統計表

| | | #1 |
|--------|---------|---------|
| 最小 T | | -23.58 |
| 最小 t | | -23.58 |
| 最大值 T | | -19.15 |
| 最大值 t | | -19.15 |
| 平均值 T | | -21.004 |
| 平均值 t | | -21.004 |
| 標準偏差 T | | 0.813 |
| 標準偏差 t | | 0.813 |
| MKT T | | -21.007 |
| MKT t | | -21.007 |
| | T: 整個週期 | |
| | t: 顯示週期 | |

Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表


FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-310; Serial No.: 1401 2233; Channel No.: 1

FREEZER TEMPERATURE RANGE : -15~-25°C

DATE: 24, Feb2022 ~ 02, Mar 2022

KIMO 14012233



統計表

| #1 | |
|--------|---------|
| 最小 T | -23.51 |
| 最小 t | -23.51 |
| 最大值 T | -19.31 |
| 最大值 t | -19.31 |
| 平均值 T | -21.069 |
| 平均值 t | -21.069 |
| 標準偏差 T | 0.86 |
| 標準偏差 t | 0.86 |
| MKT T | -21.072 |
| MKT t | -21.072 |

T: 整個週期
t: 顯示週期

Comment:


長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-310; Serial No.: 1401 2233; Channel No.: 1

FREEZER TEMPERATURE RANGE : -15~-25°C

DATE: 02, Mar 2022 ~ 24, Mar 2022



統計表

| #1 | |
|--------|---------|
| 最小 T | -23.33 |
| 最小 t | -23.33 |
| 最大值 T | -18.59 |
| 最大值 t | -18.59 |
| 平均值 T | -20.721 |
| 平均值 t | -20.721 |
| 標準偏差 T | 0.802 |
| 標準偏差 t | 0.802 |
| MKT T | -20.724 |
| MKT t | -20.724 |

T: 整個週期
t: 顯示週期

Comment:本溫度記錄器(14012233)於2022年3月2日面板顯示沒電，以備品溫度記錄器(13051668)暫代。

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE: 24, Mar 2022 ~ 31, Mar 2022



2K210322157

— v1 [°C]
Min Max Avg Sd MKT
-23.3 -19.3 -20.91 0.86 -20.85



Comment:2022年3月24日換新品溫度記錄器(2K210322157)

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE: 31, Mar 2022 ~ 29, Apr 2022



2K210322157

| | | | | |
|-----------|-------|--------|------|--------|
| — v1 [°C] | | | | |
| Min | Max | Avg | Sd | MKT |
| -23.4 | -19.1 | -20.88 | 0.86 | -20.82 |



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE: 29, Apr 2022 ~ 30, May 2022



2K210322157

— v1 [°C]
Min Max Avg Sd MKT
-23.5 -18.8 -20.83 0.86 -20.78



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE: 30, May 2022 ~ 28, Jun 2022



2K210322157

— v1 [°C]
Min Max Avg Sd MKT
-23.4 -19.1 -20.85 0.86 -20.79



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1

FREEZER TEMPERATURE RANGE : -15~-25°C


DATE: 28, Jun 2022 ~ 07, Jul 2022



2K 210322157

— v1 [°C]

| Min | Max | Avg | Sd | MKT |
|-------|-------|--------|------|--------|
| -23.3 | -15.7 | -20.91 | 0.88 | -20.85 |



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1

FREEZER TEMPERATURE RANGE : -15~-25°C


DATE: 07, Jul 2022 ~ 29, Jul 2022



2K 210322157

— v1 [°C]

| Min | Max | Avg | Sd | MKT |
|-------|-------|--------|------|--------|
| -23.5 | -17.4 | -20.98 | 0.88 | -20.92 |



Comment: 於2022年7月7日更換溫度記錄器的感溫線，調整感溫探頭後重啟。

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 29, Jul 2022 ~ 31, Aug 2022



2K 210322157

— v1 2K210322157 [°C]
Min Max Avg Sd MKT
-23.5 -16.5 -20.86 0.85 -20.81



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 31, Aug 2022 ~ 30, Sep 2022



2K 210322157

— v1 2K210322157 [°C]
Min Max Avg Sd MKT
-22.8 -17.4 -20.62 0.77 -20.58



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 30, Sep 2022 ~ 18, Oct 2022



2K 210322157

— v1 2K210322157 [°C]
Min Max Avg Sd MKT
-23.1 -19.1 -20.80 0.78 -20.76



Comment: 於2022年10月18日13:30除霜，藥品移至同冷凍櫃之上層。

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 19, Oct 2022 ~ 28, Oct 2022



2K 210322157

— v1 2K210322157 [°C]
Min Max Avg Sd MKT
-22.8 -18.0 -20.80 0.92 -20.74



Comment: 於2022年10月19日除霜完畢並於14:31重啟，並將藥品從上層移回。

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 28, Oct 2022 ~ 30, Nov 2022



2K 210322157

| | | | | | |
|-------|---------|--------|------|--------|--|
| — | v1 [°C] | | | | |
| Min | Max | Avg | Sd | MKT | |
| -23.2 | -19.1 | -20.86 | 0.92 | -20.80 | |



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 30, Nov 2022 ~ 03, Jan 2023



2K 210322157

— v1 [°C]
Min Max Avg Sd MKT
-23.2 -18.7 -21.06 0.87 -21.00



Comment: 本溫度記錄器(2K210322157)於2023年1月3日送校，期間以備品溫度記錄器(1D220605898)暫代。