

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 1D220605898; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 03, Jan 2023 ~ 30, Jan 2023



1D220605898

— v1 [°C]
Min Max Avg Sd MKT
-23.5 -18.8 -20.91 0.91 -20.85



Comment: 因下載溫度用之筆電的時區有誤，此月份溫度記錄顯示之開始與結束時間比實際時間早8hr。

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 1D220605898; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 30, Jan 2023 ~ 17, Feb 2023



1D220605898

— v1 [°C]
Min Max Avg Sd MKT
-23.5 -18.5 -20.91 0.90 -20.86



Comment: 於2023年2月17日08:50除霜，藥品移至另一冷凍櫃（冷凍1）。

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 20, Feb 2023 ~ 01, Mar 2023



2K 210322157

— v1 [°C]
Min Max Avg Sd MKT
-22.5 -19.6 -20.82 0.74 -20.78



Comment: 於2023年2月20日除霜完畢並於08:57重啟，並將藥品於09:00從冷凍1移回。本溫度記錄器 (2K210322157)完成校正，於2023年2月20日換回使用。

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 01, Mar 2023 ~ 31, Mar 2023



2K 210322157

— v1 2K 210322157 [°C]
Min Max Avg Sd MKT
-22.5 -19.1 -20.82 0.72 -20.78



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 31, Mar 2023 ~ 28, Apr 2023



2K210322157

— v1 [°C]
Min Max Avg Sd MKT
-22.6 -19.5 -20.69 0.71 -20.65



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 28, Apr 2023 ~ 29, May 2023



2K210322157

— v1 [°C]
Min Max Avg Sd MKT
-22.6 -19.2 -20.70 0.72 -20.66



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 29, May 2023 ~ 30, Jun 2023



2K210322157

— v1 [°C]
Min Max Avg Sd MKT
-22.8 -19.4 -20.82 0.73 -20.78



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 30, Jun 2023 ~ 25, Jul 2023



2K210322157

— v1 [°C]
Min Max Avg Sd MKT
-23.0 -19.3 -21.10 0.68 -21.06



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1

FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 25, Jul 2023 ~ 31, Jul 2023



2K210322157

v1 [°C]
Min Max Avg Sd MKT
-22.7 -19.3 -20.91 0.72 -20.87



Comment: 本冷凍櫃（冷凍2下層）於2023年7月25日09:48因應長庚溫控網校正，移動溫度記錄器的感溫線，故藥品暫移至冷凍2上層，並於同日11:45將藥品從上層移回。

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 31, Jul 2023 ~ 31, Aug 2023



2K210322157

v1 [°C]				
Min	Max	Avg	Sd	MKT
-22.8	-18.2	-20.89	0.72	-20.85



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 31, Aug 2023 ~ 28, Sep 2023



2K210322157

— v1 [°C]
Min Max Avg Sd MKT
-23.0 -18.6 -20.87 0.78 -20.83



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 28, Sep 2023 ~ 31, Oct 2023



2K210322157

v1 [°C]				
Min	Max	Avg	Sd	MKT
-23.1	-19.4	-20.95	0.78	-20.90



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1


FREEZER TEMPERATURE RANGE : -15~-25°C

DATE : 31, Oct 2023 ~ 30, Nov 2023



2K210322157

— v1 [°C]
Min Max Avg Sd MKT
-23.2 -18.7 -21.16 0.74 -21.12



Comment:

長庚紀念醫院—林口院區 藥品保存溫度記錄表

FREEZER : Manufacturer; SANYO; Model No.: MDF-U537D; Serial No. : 10040043; Chamber: bottom

THERMOMETER : KIMO; Model No.: KTT-220-O; Serial No.: 2K210322157; Channel No.: 1

FREEZER TEMPERATURE RANGE : -15~-25°C


DATE : 30, Nov 2023 ~ 29, Dec 2023



2K210322157

— v1 [°C]

Min	Max	Avg	Sd	MKT
-23.1	-19.3	-20.94	0.72	-20.91



Comment: