

# 長庚紀念醫院—林口院區 藥品保存溫度記錄表

PHARMACY : B1 Medical Building , Clinical Research Pharmacy.

REFRIGERATOR : Manufacturer : PHCbi ; Model No. : MPF-DU302VX-PA; Serial No. : 19070046

THERMOMETER : Manufacturer : KIMO; Model No.: KTT-220-O ; Serial No. : 2K190114208; Channel No. : 1


REFRIGERATOR TEMPERATURE RANGE : -80± 10 °C

DATE : 29, Dec 2023 ~ 31, Jan 2024



2K 190114208

— v1 [°C]  
Min Max Avg Sd MKT  
-81.9 -77.9 -80.37 0.84 -80.28



Comment:

# 長庚紀念醫院—林口院區 藥品保存溫度記錄表

PHARMACY : B1 Medical Building , Clinical Research Pharmacy.

REFRIGERATOR : Manufacturer : PHCbi ; Model No. : MPF-DU302VX-PA; Serial No. : 19070046

THERMOMETER : Manufacturer : KIMO; Model No.: KTT-220-O ; Serial No. : 2K190114208; Channel No. : 1


REFRIGERATOR TEMPERATURE RANGE : -80± 10 °C

DATE : 31, Jan 2024 ~ 28, Feb 2024



**2K 190114208**

— v1 [°C]  
Min Max Avg Sd MKT  
-81.9 -77.1 -80.41 0.86 -80.31



Comment:

# 長庚紀念醫院—林口院區 藥品保存溫度記錄表

PHARMACY : B1 Medical Building , Clinical Research Pharmacy.

REFRIGERATOR : Manufacturer : PHCbi ; Model No. : MPF-DU302VX-PA; Serial No. : 19070046

THERMOMETER : Manufacturer : KIMO; Model No.: KTT-220-O ; Serial No. : 2K190114208; Channel No. : 1

REFRIGERATOR TEMPERATURE RANGE : -80±10 °C


DATE : 28, Feb 2024 ~ 1, Apr 2024



2K 190114208

— v1 [°C]

Min	Max	Avg	Sd	MKT
-81.9	-78.7	-80.42	0.84	-80.33



Comment:

# 長庚紀念醫院—林口院區 藥品保存溫度記錄表

PHARMACY : B1 Medical Building , Clinical Research Pharmacy.

REFRIGERATOR : Manufacturer : PHCbi ; Model No. : MPF-DU302VX-PA; Serial No. : 19070046

THERMOMETER : Manufacturer : KIMO; Model No.: KTT-220-O ; Serial No. : 2K190114208; Channel No. : 1


REFRIGERATOR TEMPERATURE RANGE : -80±10 °C

DATE : 1, Apr 2024 ~ 29, Apr 2024



2K 190114208

—	v1 [°C]				
Min	Max	Avg	Sd	MKT	
-81.7	-78.9	-80.37	0.86	-80.28	



Comment: