

# 長庚紀念醫院—林口院區 藥品保存溫度記錄表

PHARMACY : B1 Medical Building , Clinical Research Pharmacy.

REFRIGERATOR : Manufacturer: PANASONIC; Model No.: MPR-1014R-PT; Serial No. : 1207 0022

THERMOMETER : Manufacturer : KIMO; Model No.: KTT-220-O; Serial No.: 2K190616069; Channel No.: 1


REFRIGERATOR TEMPERATURE RANGE : 2-8°C

DATE : 29, Dec 2023 ~ 31, Jan 2024



2K190616069

v1 [°C]				
Min	Max	Avg	Sd	MKT
3.8	6.2	4.56	0.28	4.57



Comment:

# 長庚紀念醫院—林口院區 藥品保存溫度記錄表

PHARMACY : B1 Medical Building , Clinical Research Pharmacy.

REFRIGERATOR : Manufacturer: PANASONIC; Model No.: MPR-1014R-PT; Serial No. : 1207 0022

THERMOMETER : Manufacturer : KIMO; Model No.: KTT-220-O; Serial No.: 2K190616069; Channel No.: 1


REFRIGERATOR TEMPERATURE RANGE : 2-8°C

DATE : 31, Jan 2024 ~ 28, Feb 2024



2K190616069

—	v1 [°C]				
Min	Max	Avg	Sd	MKT	
3.9	5.9	4.61	0.30	4.62	



Comment:

# 長庚紀念醫院—林口院區 藥品保存溫度記錄表

PHARMACY : B1 Medical Building , Clinical Research Pharmacy.

REFRIGERATOR : Manufacturer: PANASONIC; Model No.: MPR-1014R-PT; Serial No. : 1207 0022

THERMOMETER : Manufacturer : KIMO; Model No.: KTT-220-O; Serial No.: 2K190616069; Channel No.: 1


REFRIGERATOR TEMPERATURE RANGE : 2-8°C

DATE : 28, Feb 2024 ~ 1, Apr 2024



2K190616069

— v1 [°C]				
Min	Max	Avg	Sd	MKT
3.6	6.7	4.94	0.32	4.95



Comment:

# 長庚紀念醫院—林口院區 藥品保存溫度記錄表

PHARMACY : B1 Medical Building , Clinical Research Pharmacy.

REFRIGERATOR : Manufacturer: PANASONIC; Model No.: MPR-1014R-PT; Serial No. : 1207 0022

THERMOMETER : Manufacturer : KIMO; Model No.: KTT-220-O; Serial No.: 2K190616069; Channel No.: 1


REFRIGERATOR TEMPERATURE RANGE : 2-8°C

DATE : 1, Apr 2024 ~ 29, Apr 2024



2K190616069

v1 [°C]				
Min	Max	Avg	Sd	MKT
4.2	6.0	4.86	0.30	4.87



Comment: